

MAULDEN PARISH COUNCIL
NON CONFIDENTIAL
COMMITTEE: Maulden Parish Council
DATE: 09/04/2026
REPORT BY: Sean Rudkin
SUBJECT: Defibrillator report for April 2026

I conducted a check on the parish council owned defibrillators at the Village Hall, Budgens, Hilsons and "Pinky's" on 9th April 2026. In all cases, the equipment was found to be complete and in working order with exceptions noted below. Comprehensive prep kit was available in all cabinets and integrated cabinet lights working.

- The Hilsons unit is warning that the 9v battery needs changing
- Replacement pads are still required for Hilsons, Pinky's and Budgens

On the 15th March, the Ambulance service reported that the defib at Budgens may have been accessed. This was checked and the ok for return to service given on the 18th March.

On the 20th March, the Ambulance service reported that the defib at the Village Hall may have been accessed. This was checked and the ok for return to service given on the same day.